

CERTIFICATE No.02638

CUSTOMER DETAILS	
NAME	GILLETT ENVIRONMENTAL
ADDRESS	BOUNDARY ROAD LYTHAM
POST CODE	FY8 5HU

MACHINE DETAILS	
INDICATOR MANUF/TYPE	VISHAY / VT300A
INDICATOR SERIAL No	T10310404
SCALE TYPE	WEIGHBRIDGE
BASE WORK MANUF/No	- / -
SOFTWARE VERSION	-
CAPACITY	35,000KG
INCREMENTS	20KG
LOCATION	AS ABOVE

CALIBRATION DETAILS	
NEXT CERTIFICATE DUE	01.2023
DATE OF CALIBRATION	27.01.2022
TEST WEIGHT CERT NO	9017
ENGINEERS NAME	BEN POWELL
SIGNATURE	<i>Andrew D Scaville</i>


LINEARITY AND HYSTERESIS TEST						
AS FOUND						
	NO TARE		TARE=		TARE=	
	INDICATION		INDICATION		INDICATION	
LOAD	LOADING	UNLOADING	LOADING	UNLOADING	LOADING	UNLOADING
0 KG	0 KG					
2,000 KG	2,000 KG					
10,000 KG	10,020 KG					
20,000 KG	20,020 KG					
24,260 KG	24,280 KG					
29,220 KG	29,240 KG					

DEFINITIVE TEST						
	NO TARE		TARE=		TARE=	
	INDICATION		INDICATION		INDICATION	
LOAD	LOADING	UNLOADING	LOADING	UNLOADING	LOADING	UNLOADING

REPEATABILITY TEST						
	AS FOUND			DEFINITIVE		
LOAD	IND	IND	IND	IND	IND	IND

DISCRIMINATION/SENSITIVITY TEST			
LOAD	INDICATION	ADDED	CHANGE

ECCENTRIC LOAD TEST				
	AS FOUND		DEFINITIVE	
POSITION	LOAD	IND	LOAD	IND
1	6,260 KG	6,260 KG		
2	6,260 KG	6,260 KG		
3	6,260 KG	6,260 KG		
4	6,260 KG	6,260 KG		
5				
6				
7				
8				



CALIBRATION CARRIED OUT IN ACCORDANCE WITH THE U.K.W.F. CODE OF PRACTICE

SCALES SPARES & SERVICES LTD
 THORNHILL ROAD BUSINESS PARK, TENTER FIELDS, DEWSBURY, WF12 9QT
 TELEPHONE: 01924 464967

MEMBERS OF THE UK WEIGHING FEDERATION